

Automated External Defibrillator

Inspection Tag

- Check the ready indicator is not flashing and device is not emitting an audible alert
- \blacksquare Check general condition of the device

\blacksquare Check electrodes are in date and battery charged

Date	Checked by	Status indicator OK (Y / N)	Pads in date (Y / N)	Battery charged (Y / N)
Do ne	ot remove this tag	©Act Fast Clinical 2023		V1.0



Automated External Defibrillator

Inspection Tag

- Check the ready indicator is not flashing and device is not emitting an audible alert
- \blacksquare Check general condition of the device

 \blacksquare Check electrodes are in date and battery charged

Date	Checked by	Status indicator OK (Y / N)	Pads in date (Y / N)	Battery charged (Y / N)
Do not remove this tag ©Act Fast Clinical 20				V1.0



Automated External Defibrillator

Inspection Tag

- Check the ready indicator is not flashing and device is not emitting an audible alert
- \blacksquare Check general condition of the device
- \blacksquare Check electrodes are in date and battery charged

Date	Checked by	Status indicator OK (Y / N)	Pads in date (Y / N)	Battery charged (Y / N)
Dono	ot remove this tag	©Act Fast Clinical 2023		V1.0



Automated External Defibrillator

Inspection Tag

- Check the ready indicator is not flashing and device is not emitting an audible alert
- \blacksquare Check general condition of the device
- ☑ Check electrodes are in date and battery charged

Date	Checked by	Status indicator OK (Y / N)	Pads in date (Y / N)	Battery charged (Y / N)
Do n	ot remove this tag	©Act Fast Clinical 2023		V1.0







